

# Certificate



Certificate no.

CU 72180492 01

**License Holder:**

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122  
USA

**Manufacturing Plant:**

Illumina Singapore Pte. Ltd.  
North Tech Lobby 3 #02-13/18  
29 Woodlands Industrial Park E1  
Singapore 757716  
Singapore

Test report no.: USA-JR 31880498 002

Client Reference: Carol Escano

**Tested to:**

UL 61010-1:2012 R4.16  
UL 61010-2-010:2015  
UL 61010-2-081:2015  
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2  
CAN/CSA-C22.2 NO. 61010-2-010:15  
CAN/CSA-C22.2 NO. 61010-2-081:15

**Certified Product:** Next Generation Sequencer

**License Fee - Units**

Model Designation: iSeq 100

7

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 80W

Protection Class: I

Appendix: 1, 1-4

7

Licensed Test mark:

Date of Issue

(day/mo/yr)

16/03/2018

